

FBI  
997904

## MEMORANDUM

TO: File

FROM: [REDACTED]

STATINTL

DATE: 29 October 1965

SUBJECT: Subjects discussed with customer on Project [REDACTED] during his visit on 25 October 1965.

STATINTL

(1) Film Hold Down Bars-tables 2 & 3. During the design of the transport for the microscopes, it became apparent that the carriage will pass over these bars when it is used or parked at the end of the table. If the PI begins to crank the film, the bars would be raised and strike the bottom of the carriage. We suggested eliminating the film hold down bars on tables 2 & 3 because of this problem and also because the film must be held flat with a glass to make measurements. A better solution would be to add an automatic glass film hold down plate as described in [REDACTED] proposal [REDACTED]

STATINTL

STATINTL

The customer agreed, but said it would be difficult to change the contract. Thus we shall incorporate a microswitch near the XX carriage bar which will disengage the film drive mechanism when the carriage is over the film hold down bar.

STATINTL

(2) The microscope carriage and [REDACTED] table were reviewed and found satisfactory. The fluorescent tracking light sources were reviewed and the design was acceptable to the customer.

(3) A tentative schedule for the customer to review the progress on the light tables was set for the first week in December. A phone contact will be made the latter part of November to be certain there will be hardware available, and thus assure the trip worth while.

Declass Review by NIMA / DoD